



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/09/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 10:55 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.72       | 5.72      | 5.62      | 5.82  | 0.00       |
| Aroclor-1016-2       | (2) | 5.74       | 5.74      | 5.64      | 5.84  | 0.00       |
| Aroclor-1016-3       | (3) | 5.81       | 5.81      | 5.71      | 5.91  | 0.01       |
| Aroclor-1016-4       | (4) | 5.91       | 5.91      | 5.81      | 6.01  | 0.00       |
| Aroclor-1016-5       | (5) | 6.20       | 6.20      | 6.10      | 6.30  | 0.00       |
| Aroclor-1260-1       | (1) | 7.34       | 7.34      | 7.24      | 7.44  | 0.00       |
| Aroclor-1260-2       | (2) | 7.60       | 7.60      | 7.50      | 7.70  | 0.00       |
| Aroclor-1260-3       | (3) | 7.97       | 7.97      | 7.87      | 8.07  | 0.00       |
| Aroclor-1260-4       | (4) | 8.20       | 8.20      | 8.10      | 8.30  | 0.00       |
| Aroclor-1260-5       | (5) | 8.53       | 8.53      | 8.43      | 8.63  | 0.00       |
| Tetrachloro-m-xylene |     | 4.53       | 4.53      | 4.43      | 4.63  | 0.00       |
| Decachlorobiphenyl   |     | 10.47      | 10.46     | 10.36     | 10.56 | -0.01      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/09/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 10:55 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.78       | 4.79      | 4.69      | 4.89 | 0.01       |
| Aroclor-1016-2       | (2) | 4.80       | 4.80      | 4.70      | 4.90 | 0.00       |
| Aroclor-1016-3       | (3) | 4.98       | 4.98      | 4.88      | 5.08 | 0.00       |
| Aroclor-1016-4       | (4) | 5.02       | 5.02      | 4.92      | 5.12 | 0.00       |
| Aroclor-1016-5       | (5) | 5.23       | 5.24      | 5.14      | 5.34 | 0.01       |
| Aroclor-1260-1       | (1) | 6.28       | 6.28      | 6.18      | 6.38 | 0.00       |
| Aroclor-1260-2       | (2) | 6.47       | 6.47      | 6.37      | 6.57 | 0.00       |
| Aroclor-1260-3       | (3) | 6.62       | 6.63      | 6.53      | 6.73 | 0.01       |
| Aroclor-1260-4       | (4) | 7.10       | 7.10      | 7.00      | 7.20 | 0.01       |
| Aroclor-1260-5       | (5) | 7.34       | 7.34      | 7.24      | 7.44 | 0.00       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   |     | 8.75       | 8.75      | 8.65      | 8.85 | 0.00       |