



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 09:39 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.72	5.72	5.62	5.82	0.00
Aroclor-1016-2	(2)	5.74	5.74	5.64	5.84	0.00
Aroclor-1016-3	(3)	5.81	5.81	5.71	5.91	0.01
Aroclor-1016-4	(4)	5.91	5.91	5.81	6.01	0.00
Aroclor-1016-5	(5)	6.20	6.20	6.10	6.30	0.00
Aroclor-1260-1	(1)	7.34	7.34	7.24	7.44	0.00
Aroclor-1260-2	(2)	7.60	7.60	7.50	7.70	0.00
Aroclor-1260-3	(3)	7.97	7.97	7.87	8.07	0.00
Aroclor-1260-4	(4)	8.20	8.20	8.10	8.30	0.00
Aroclor-1260-5	(5)	8.53	8.53	8.43	8.63	0.00
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.46	10.46	10.36	10.56	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 09:39 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-4	(4)	5.02	5.02	4.92	5.12	0.00
Aroclor-1016-5	(5)	5.23	5.24	5.14	5.34	0.01
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.46	6.47	6.37	6.57	0.01
Aroclor-1260-3	(3)	6.62	6.63	6.53	6.73	0.01
Aroclor-1260-4	(4)	7.10	7.10	7.00	7.20	0.01
Aroclor-1260-5	(5)	7.34	7.34	7.24	7.44	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.75	8.75	8.65	8.85	0.00