



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 16:30 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.72       | 5.72      | 5.62      | 5.82  | 0.00       |
| Aroclor-1242-2 (2)   | 5.74       | 5.74      | 5.64      | 5.84  | 0.00       |
| Aroclor-1242-3 (3)   | 5.81       | 5.81      | 5.71      | 5.91  | 0.00       |
| Aroclor-1242-4 (4)   | 5.91       | 5.91      | 5.81      | 6.01  | 0.00       |
| Aroclor-1242-5 (5)   | 6.65       | 6.65      | 6.55      | 6.75  | 0.00       |
| Tetrachloro-m-xylene | 4.53       | 4.53      | 4.43      | 4.63  | 0.00       |
| Decachlorobiphenyl   | 10.47      | 10.46     | 10.36     | 10.56 | -0.01      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 16:30 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1242-2 (2)   | 4.80       | 4.80      | 4.70      | 4.90 | 0.00       |
| Aroclor-1242-3 (3)   | 4.98       | 4.98      | 4.88      | 5.08 | 0.00       |
| Aroclor-1242-4 (4)   | 5.06       | 5.07      | 4.97      | 5.17 | 0.01       |
| Aroclor-1242-5 (5)   | 5.59       | 5.59      | 5.49      | 5.69 | 0.00       |
| Tetrachloro-m-xylene | 3.68       | 3.69      | 3.59      | 3.79 | 0.01       |
| Decachlorobiphenyl   | 8.75       | 8.75      | 8.65      | 8.85 | 0.00       |