



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 17:02 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1254-1 (1)   | 6.59       | 6.59      | 6.49      | 6.69  | 0.00       |
| Aroclor-1254-2 (2)   | 6.81       | 6.81      | 6.71      | 6.91  | 0.00       |
| Aroclor-1254-3 (3)   | 7.18       | 7.18      | 7.08      | 7.28  | 0.00       |
| Aroclor-1254-4 (4)   | 7.47       | 7.47      | 7.37      | 7.57  | 0.00       |
| Aroclor-1254-5 (5)   | 7.89       | 7.89      | 7.79      | 7.99  | 0.00       |
| Tetrachloro-m-xylene | 4.53       | 4.53      | 4.43      | 4.63  | 0.00       |
| Decachlorobiphenyl   | 10.47      | 10.45     | 10.35     | 10.55 | -0.02      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 17:02 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.59       | 5.59      | 5.49      | 5.69 | 0.00       |
| Aroclor-1254-2 (2)   | 5.74       | 5.74      | 5.64      | 5.84 | 0.00       |
| Aroclor-1254-3 (3)   | 6.15       | 6.15      | 6.05      | 6.25 | 0.00       |
| Aroclor-1254-4 (4)   | 6.38       | 6.38      | 6.28      | 6.48 | 0.00       |
| Aroclor-1254-5 (5)   | 6.80       | 6.80      | 6.70      | 6.90 | 0.00       |
| Tetrachloro-m-xylene | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   | 8.75       | 8.75      | 8.65      | 8.85 | 0.00       |