



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/11/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 01:49 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.72	5.72	5.62	5.82	0.00
Aroclor-1248-2 (2)	6.00	6.00	5.90	6.10	0.00
Aroclor-1248-3 (3)	6.21	6.20	6.10	6.30	-0.01
Aroclor-1248-4 (4)	6.62	6.61	6.51	6.71	-0.01
Aroclor-1248-5 (5)	6.66	6.65	6.55	6.75	-0.01
Tetrachloro-m-xylene	4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl	10.47	10.45	10.35	10.55	-0.02



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/11/2023 Initial Calibration Date(s): 09/26/2023 09/27/2023

Continuing Calib Time: 01:49 Initial Calibration Time(s): 16:50 00:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1248-2 (2)	5.02	5.02	4.92	5.12	0.00
Aroclor-1248-3 (3)	5.06	5.07	4.97	5.17	0.01
Aroclor-1248-4 (4)	5.24	5.24	5.14	5.34	0.00
Aroclor-1248-5 (5)	5.63	5.64	5.54	5.74	0.01
Tetrachloro-m-xylene	3.69	3.68	3.58	3.78	0.00
Decachlorobiphenyl	8.75	8.75	8.65	8.85	0.00