



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/16/2023 Initial Calibration Date(s): 10/13/2023 10/13/2023

Continuing Calib Time: 14:32 Initial Calibration Time(s): 15:32 22:55

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.74	5.73	5.63	5.83	-0.01
Aroclor-1016-1	(1)	5.74	5.74	5.64	5.84	0.00
Aroclor-1016-2	(2)	5.77	5.75	5.65	5.85	-0.02
Aroclor-1016-2	(2)	5.77	5.76	5.66	5.86	-0.01
Aroclor-1016-3	(3)	5.83	5.81	5.71	5.91	-0.02
Aroclor-1016-3	(3)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-4	(4)	5.93	5.91	5.81	6.01	-0.02
Aroclor-1016-4	(4)	5.93	5.93	5.83	6.03	0.00
Aroclor-1016-5	(5)	6.23	6.21	6.11	6.31	-0.02
Aroclor-1016-5	(5)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-1	(1)	7.37	7.35	7.25	7.45	-0.02
Aroclor-1260-1	(1)	7.37	7.37	7.27	7.47	0.00
Aroclor-1260-2	(2)	7.63	7.61	7.51	7.71	-0.02
Aroclor-1260-2	(2)	7.63	7.63	7.53	7.73	0.00
Aroclor-1260-3	(3)	8.00	7.98	7.88	8.08	-0.01
Aroclor-1260-3	(3)	8.00	7.99	7.89	8.09	0.00
Aroclor-1260-4	(4)	8.23	8.21	8.11	8.31	-0.02
Aroclor-1260-4	(4)	8.23	8.22	8.12	8.32	-0.01
Aroclor-1260-5	(5)	8.57	8.54	8.44	8.64	-0.03
Aroclor-1260-5	(5)	8.57	8.56	8.46	8.66	-0.01
Tetrachloro-m-xylene		4.55	4.54	4.44	4.64	-0.01
Decachlorobiphenyl		10.52	10.48	10.38	10.58	-0.04
Decachlorobiphenyl		10.52	10.51	10.41	10.61	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/16/2023 Initial Calibration Date(s): 10/13/2023 10/13/2023

Continuing Calib Time: 14:32 Initial Calibration Time(s): 15:32 22:55

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-3	(3)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-4	(4)	5.03	5.03	4.93	5.13	0.00
Aroclor-1016-5	(5)	5.24	5.24	5.14	5.34	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-2	(2)	6.47	6.47	6.37	6.57	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	7.11	7.10	7.00	7.20	-0.01
Aroclor-1260-5	(5)	7.35	7.34	7.24	7.44	-0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.76	8.75	8.65	8.85	-0.01
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00