



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/19/2023 09/20/2023

Continuing Calib Time: 09:04 Initial Calibration Time(s): 21:37 04:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.58	4.58	4.48	4.68	0.01
Aroclor-1016-2	(2)	4.60	4.60	4.50	4.70	0.01
Aroclor-1016-3	(3)	4.65	4.66	4.56	4.76	0.01
Aroclor-1016-4	(4)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-5	(5)	5.04	5.04	4.94	5.14	0.01
Aroclor-1260-1	(1)	6.15	6.15	6.05	6.25	0.00
Aroclor-1260-2	(2)	6.41	6.41	6.31	6.51	0.00
Aroclor-1260-3	(3)	6.76	6.77	6.67	6.87	0.01
Aroclor-1260-4	(4)	6.98	6.99	6.89	7.09	0.01
Aroclor-1260-5	(5)	7.30	7.30	7.20	7.40	0.00
Tetrachloro-m-xylene		3.46	3.46	3.36	3.56	0.00
Decachlorobiphenyl		8.70	8.70	8.60	8.80	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/19/2023 09/20/2023

Continuing Calib Time: 09:04 Initial Calibration Time(s): 21:37 04:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-2	(2)	3.81	3.81	3.71	3.91	0.00
Aroclor-1016-3	(3)	3.97	3.97	3.87	4.07	0.00
Aroclor-1016-4	(4)	4.02	4.02	3.92	4.12	0.00
Aroclor-1016-5	(5)	4.21	4.22	4.12	4.32	0.01
Aroclor-1260-1	(1)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-2	(2)	5.40	5.41	5.31	5.51	0.01
Aroclor-1260-3	(3)	5.55	5.55	5.45	5.65	0.00
Aroclor-1260-4	(4)	6.01	6.01	5.91	6.11	0.00
Aroclor-1260-5	(5)	6.26	6.26	6.16	6.36	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.58	7.59	7.49	7.69	0.01