



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 10/10/2023 10/10/2023

Continuing Calib Time: 16:32 Initial Calibration Time(s): 10:41 13:22

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-2	(2)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.17	5.17	5.07	5.27	0.00
Aroclor-1016-5	(5)	5.49	5.49	5.39	5.59	0.00
Aroclor-1260-1	(1)	6.72	6.72	6.62	6.82	0.00
Aroclor-1260-2	(2)	7.00	7.00	6.90	7.10	0.00
Aroclor-1260-3	(3)	7.39	7.40	7.30	7.50	0.01
Aroclor-1260-4	(4)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-5	(5)	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene		3.74	3.74	3.64	3.84	0.00
Decachlorobiphenyl		9.76	9.77	9.67	9.87	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 10/10/2023 10/10/2023

Continuing Calib Time: 16:32 Initial Calibration Time(s): 10:41 13:22

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.18	4.18	4.08	4.28	0.00
Aroclor-1016-2 (2)	4.20	4.20	4.10	4.30	0.00
Aroclor-1016-3 (3)	4.38	4.38	4.28	4.48	0.00
Aroclor-1016-4 (4)	4.43	4.44	4.34	4.54	0.01
Aroclor-1016-5 (5)	4.66	4.66	4.56	4.76	0.00
Aroclor-1260-1 (1)	5.78	5.78	5.68	5.88	0.00
Aroclor-1260-2 (2)	5.99	5.99	5.89	6.09	0.00
Aroclor-1260-3 (3)	6.15	6.15	6.05	6.25	0.00
Aroclor-1260-4 (4)	6.67	6.67	6.57	6.77	0.00
Aroclor-1260-5 (5)	6.94	6.94	6.84	7.04	0.00
Tetrachloro-m-xylene	3.03	3.03	2.93	3.13	0.00
Decachlorobiphenyl	8.44	8.44	8.34	8.54	0.00