



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/27/2023 09/27/2023

Client Sample No.: CCAL01 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PO098624.D Time Analyzed: 08:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.654	5.558	5.758	491.610	500.000	-1.7
Aroclor-1016-2	5.677	5.581	5.781	492.610	500.000	-1.5
Aroclor-1016-3	5.739	5.643	5.843	509.110	500.000	1.8
Aroclor-1016-4	5.839	5.742	5.942	506.420	500.000	1.3
Aroclor-1016-5	6.135	6.038	6.238	495.240	500.000	-1.0
Aroclor-1260-1	7.264	7.167	7.367	501.260	500.000	0.3
Aroclor-1260-2	7.521	7.425	7.625	488.630	500.000	-2.3
Aroclor-1260-3	7.882	7.785	7.985	521.720	500.000	4.3
Aroclor-1260-4	8.109	8.012	8.212	537.680	500.000	7.5
Aroclor-1260-5	8.435	8.338	8.538	484.060	500.000	-3.2
Decachlorobiphenyl	10.291	10.194	10.394	49.040	50.000	-1.9
Tetrachloro-m-xylene	4.479	4.385	4.585	49.720	50.000	-0.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/27/2023 09/27/2023

Client Sample No.: CCAL01 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PO098624.D Time Analyzed: 08:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.623	4.823	493.590	500.000	-1.3
Aroclor-1016-2	4.736	4.642	4.842	487.160	500.000	-2.6
Aroclor-1016-3	4.912	4.817	5.017	512.800	500.000	2.6
Aroclor-1016-4	4.954	4.860	5.060	526.110	500.000	5.2
Aroclor-1016-5	5.168	5.074	5.274	527.730	500.000	5.5
Aroclor-1260-1	6.201	6.109	6.309	496.020	500.000	-0.8
Aroclor-1260-2	6.390	6.297	6.497	492.090	500.000	-1.6
Aroclor-1260-3	6.544	6.451	6.651	493.010	500.000	-1.4
Aroclor-1260-4	7.016	6.924	7.124	504.420	500.000	0.9
Aroclor-1260-5	7.259	7.166	7.366	476.390	500.000	-4.7
Decachlorobiphenyl	8.641	8.552	8.752	51.500	50.000	3.0
Tetrachloro-m-xylene	3.634	3.538	3.738	49.620	50.000	-0.8