



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/27/2023 09/27/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PO098626.D Time Analyzed: 09:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.655	5.558	5.758	485.080	500.000	-3.0
Aroclor-1016-2	5.678	5.581	5.781	491.450	500.000	-1.7
Aroclor-1016-3	5.741	5.643	5.843	499.370	500.000	-0.1
Aroclor-1016-4	5.839	5.742	5.942	500.770	500.000	0.2
Aroclor-1016-5	6.135	6.038	6.238	501.320	500.000	0.3
Aroclor-1260-1	7.265	7.167	7.367	503.020	500.000	0.6
Aroclor-1260-2	7.523	7.425	7.625	489.030	500.000	-2.2
Aroclor-1260-3	7.883	7.785	7.985	524.470	500.000	4.9
Aroclor-1260-4	8.110	8.012	8.212	539.480	500.000	7.9
Aroclor-1260-5	8.437	8.338	8.538	481.430	500.000	-3.7
Decachlorobiphenyl	10.293	10.194	10.394	49.140	50.000	-1.7
Tetrachloro-m-xylene	4.480	4.385	4.585	49.490	50.000	-1.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/27/2023 09/27/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PO098626.D Time Analyzed: 09:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.623	4.823	491.870	500.000	-1.6
Aroclor-1016-2	4.735	4.642	4.842	488.430	500.000	-2.3
Aroclor-1016-3	4.911	4.817	5.017	503.140	500.000	0.6
Aroclor-1016-4	4.953	4.860	5.060	517.440	500.000	3.5
Aroclor-1016-5	5.166	5.074	5.274	525.840	500.000	5.2
Aroclor-1260-1	6.202	6.109	6.309	494.120	500.000	-1.2
Aroclor-1260-2	6.390	6.297	6.497	492.570	500.000	-1.5
Aroclor-1260-3	6.544	6.451	6.651	490.010	500.000	-2.0
Aroclor-1260-4	7.016	6.924	7.124	501.290	500.000	0.3
Aroclor-1260-5	7.259	7.166	7.366	471.340	500.000	-5.7
Decachlorobiphenyl	8.643	8.552	8.752	51.680	50.000	3.4
Tetrachloro-m-xylene	3.633	3.538	3.738	50.200	50.000	0.4