



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2023 09/26/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PP060812.D Time Analyzed: 06:07

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|----------------------|--------|-----------|--------|--------------------|-------------------|-----|
|                      |        | FROM      | TO     |                    |                   |     |
| Aroclor-1016-1       | 5.719  | 5.619     | 5.819  | 525.540            | 500.000           | 5.1 |
| Aroclor-1016-2       | 5.742  | 5.642     | 5.842  | 519.820            | 500.000           | 4.0 |
| Aroclor-1016-3       | 5.804  | 5.706     | 5.906  | 525.790            | 500.000           | 5.2 |
| Aroclor-1016-4       | 5.905  | 5.806     | 6.006  | 527.870            | 500.000           | 5.6 |
| Aroclor-1016-5       | 6.204  | 6.104     | 6.304  | 521.780            | 500.000           | 4.4 |
| Aroclor-1260-1       | 7.343  | 7.244     | 7.444  | 513.790            | 500.000           | 2.8 |
| Aroclor-1260-2       | 7.602  | 7.502     | 7.702  | 508.000            | 500.000           | 1.6 |
| Aroclor-1260-3       | 7.965  | 7.865     | 8.065  | 523.370            | 500.000           | 4.7 |
| Aroclor-1260-4       | 8.197  | 8.096     | 8.296  | 522.690            | 500.000           | 4.5 |
| Aroclor-1260-5       | 8.531  | 8.430     | 8.630  | 507.380            | 500.000           | 1.5 |
| Decachlorobiphenyl   | 10.462 | 10.356    | 10.556 | 51.100             | 50.000            | 2.2 |
| Tetrachloro-m-xylene | 4.530  | 4.432     | 4.632  | 52.280             | 50.000            | 4.6 |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2023 09/26/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PP060812.D Time Analyzed: 06:07

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.779 | 4.685     | 4.885 | 494.030            | 500.000           | -1.2 |
| Aroclor-1016-2       | 4.799 | 4.704     | 4.904 | 485.470            | 500.000           | -2.9 |
| Aroclor-1016-3       | 4.977 | 4.882     | 5.082 | 493.940            | 500.000           | -1.2 |
| Aroclor-1016-4       | 5.019 | 4.924     | 5.124 | 478.860            | 500.000           | -4.2 |
| Aroclor-1016-5       | 5.234 | 5.139     | 5.339 | 451.550            | 500.000           | -9.7 |
| Aroclor-1260-1       | 6.276 | 6.182     | 6.382 | 466.530            | 500.000           | -6.7 |
| Aroclor-1260-2       | 6.464 | 6.370     | 6.570 | 465.620            | 500.000           | -6.9 |
| Aroclor-1260-3       | 6.620 | 6.525     | 6.725 | 473.060            | 500.000           | -5.4 |
| Aroclor-1260-4       | 7.095 | 7.000     | 7.200 | 482.480            | 500.000           | -3.5 |
| Aroclor-1260-5       | 7.337 | 7.242     | 7.442 | 482.100            | 500.000           | -3.6 |
| Decachlorobiphenyl   | 8.745 | 8.651     | 8.851 | 48.830             | 50.000            | -2.3 |
| Tetrachloro-m-xylene | 3.682 | 3.584     | 3.784 | 53.680             | 50.000            | 7.4  |