



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/12/2023 10/12/2023

Client Sample No.: CCAL05 Date Analyzed: 10/16/2023

Lab Sample No.: AR1660CCC500 Data File : PP060991.D Time Analyzed: 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.752	5.627	5.827	508.510	500.000	1.7
Aroclor-1016-1	5.752	5.641	5.841	508.510	500.000	1.7
Aroclor-1016-2	5.774	5.650	5.850	505.940	500.000	1.2
Aroclor-1016-2	5.774	5.664	5.864	505.940	500.000	1.2
Aroclor-1016-3	5.838	5.713	5.913	510.910	500.000	2.2
Aroclor-1016-3	5.838	5.728	5.928	510.910	500.000	2.2
Aroclor-1016-4	5.938	5.814	6.014	513.600	500.000	2.7
Aroclor-1016-4	5.938	5.828	6.028	513.600	500.000	2.7
Aroclor-1016-5	6.238	6.113	6.313	518.540	500.000	3.7
Aroclor-1016-5	6.238	6.127	6.327	518.540	500.000	3.7
Aroclor-1260-1	7.379	7.253	7.453	532.520	500.000	6.5
Aroclor-1260-1	7.379	7.268	7.468	532.520	500.000	6.5
Aroclor-1260-2	7.638	7.512	7.712	512.370	500.000	2.5
Aroclor-1260-2	7.638	7.527	7.727	512.370	500.000	2.5
Aroclor-1260-3	8.002	7.875	8.075	507.220	500.000	1.4
Aroclor-1260-3	8.002	7.890	8.090	507.220	500.000	1.4
Aroclor-1260-4	8.236	8.107	8.307	513.610	500.000	2.7
Aroclor-1260-4	8.236	8.124	8.324	513.610	500.000	2.7
Aroclor-1260-5	8.573	8.441	8.641	513.600	500.000	2.7
Aroclor-1260-5	8.573	8.460	8.660	513.600	500.000	2.7
Decachlorobiphenyl	10.527	10.379	10.579	53.380	50.000	6.8
Decachlorobiphenyl	10.527	10.407	10.607	53.380	50.000	6.8
Tetrachloro-m-xylene	4.559	4.438	4.638	51.500	50.000	3.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/12/2023 10/12/2023

Client Sample No.: CCAL05 Date Analyzed: 10/16/2023

Lab Sample No.: AR1660CCC500 Data File : PP060991.D Time Analyzed: 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.789	4.686	4.886	458.340	500.000	-8.3
Aroclor-1016-2	4.808	4.705	4.905	458.680	500.000	-8.3
Aroclor-1016-3	4.986	4.883	5.083	461.090	500.000	-7.8
Aroclor-1016-3	4.986	4.884	5.084	461.090	500.000	-7.8
Aroclor-1016-4	5.029	4.926	5.126	450.670	500.000	-9.9
Aroclor-1016-5	5.244	5.141	5.341	501.480	500.000	0.3
Aroclor-1260-1	6.287	6.182	6.382	459.030	500.000	-8.2
Aroclor-1260-1	6.287	6.183	6.383	459.030	500.000	-8.2
Aroclor-1260-2	6.477	6.371	6.571	457.320	500.000	-8.5
Aroclor-1260-2	6.477	6.372	6.572	457.320	500.000	-8.5
Aroclor-1260-3	6.631	6.526	6.726	463.040	500.000	-7.4
Aroclor-1260-4	7.107	7.002	7.202	476.270	500.000	-4.7
Aroclor-1260-5	7.350	7.244	7.444	463.590	500.000	-7.3
Decachlorobiphenyl	8.762	8.654	8.854	46.560	50.000	-6.9
Decachlorobiphenyl	8.762	8.655	8.855	46.560	50.000	-6.9
Tetrachloro-m-xylene	3.688	3.587	3.787	45.970	50.000	-8.1