



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/12/2023 10/12/2023

Client Sample No.: CCAL06 Date Analyzed: 10/16/2023

Lab Sample No.: AR1660CCC500 Data File : PP060999.D Time Analyzed: 14:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.743	5.627	5.827	523.310	500.000	4.7
Aroclor-1016-1	5.743	5.641	5.841	523.310	500.000	4.7
Aroclor-1016-2	5.767	5.650	5.850	523.170	500.000	4.6
Aroclor-1016-2	5.767	5.664	5.864	523.170	500.000	4.6
Aroclor-1016-3	5.830	5.713	5.913	527.720	500.000	5.5
Aroclor-1016-3	5.830	5.728	5.928	527.720	500.000	5.5
Aroclor-1016-4	5.930	5.814	6.014	529.760	500.000	6.0
Aroclor-1016-4	5.930	5.828	6.028	529.760	500.000	6.0
Aroclor-1016-5	6.230	6.113	6.313	536.170	500.000	7.2
Aroclor-1016-5	6.230	6.127	6.327	536.170	500.000	7.2
Aroclor-1260-1	7.372	7.253	7.453	561.290	500.000	12.3
Aroclor-1260-1	7.372	7.268	7.468	561.290	500.000	12.3
Aroclor-1260-2	7.631	7.512	7.712	536.120	500.000	7.2
Aroclor-1260-2	7.631	7.527	7.727	536.120	500.000	7.2
Aroclor-1260-3	7.995	7.875	8.075	530.600	500.000	6.1
Aroclor-1260-3	7.995	7.890	8.090	530.600	500.000	6.1
Aroclor-1260-4	8.228	8.107	8.307	528.520	500.000	5.7
Aroclor-1260-4	8.228	8.124	8.324	528.520	500.000	5.7
Aroclor-1260-5	8.566	8.441	8.641	526.390	500.000	5.3
Aroclor-1260-5	8.566	8.460	8.660	526.390	500.000	5.3
Decachlorobiphenyl	10.517	10.379	10.579	55.480	50.000	11.0
Decachlorobiphenyl	10.517	10.407	10.607	55.480	50.000	11.0
Tetrachloro-m-xylene	4.552	4.438	4.638	52.910	50.000	5.8



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/12/2023 10/12/2023

Client Sample No.: CCAL06 Date Analyzed: 10/16/2023

Lab Sample No.: AR1660CCC500 Data File : PP060999.D Time Analyzed: 14:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.787	4.686	4.886	473.790	500.000	-5.2
Aroclor-1016-2	4.806	4.705	4.905	474.730	500.000	-5.1
Aroclor-1016-3	4.984	4.883	5.083	475.650	500.000	-4.9
Aroclor-1016-3	4.984	4.884	5.084	475.650	500.000	-4.9
Aroclor-1016-4	5.026	4.926	5.126	457.270	500.000	-8.5
Aroclor-1016-5	5.242	5.141	5.341	522.230	500.000	4.4
Aroclor-1260-1	6.284	6.182	6.382	472.610	500.000	-5.5
Aroclor-1260-1	6.284	6.183	6.383	472.610	500.000	-5.5
Aroclor-1260-2	6.474	6.371	6.571	472.510	500.000	-5.5
Aroclor-1260-2	6.474	6.372	6.572	472.510	500.000	-5.5
Aroclor-1260-3	6.628	6.526	6.726	472.520	500.000	-5.5
Aroclor-1260-4	7.105	7.002	7.202	477.230	500.000	-4.6
Aroclor-1260-5	7.347	7.244	7.444	479.910	500.000	-4.0
Decachlorobiphenyl	8.760	8.654	8.854	48.240	50.000	-3.5
Decachlorobiphenyl	8.760	8.655	8.855	48.240	50.000	-3.5
Tetrachloro-m-xylene	3.688	3.587	3.787	47.530	50.000	-4.9