



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/19/2023 09/19/2023

Client Sample No.: CCAL01 Date Analyzed: 10/09/2023

Lab Sample No.: AR1660CCC500 Data File : PQ063583.D Time Analyzed: 22:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.576	4.479	4.679	485.960	500.000	-2.8
Aroclor-1016-2	4.595	4.498	4.698	479.830	500.000	-4.0
Aroclor-1016-3	4.654	4.557	4.757	483.380	500.000	-3.3
Aroclor-1016-4	4.747	4.649	4.849	490.640	500.000	-1.9
Aroclor-1016-5	5.037	4.939	5.139	492.220	500.000	-1.6
Aroclor-1260-1	6.148	6.049	6.249	519.180	500.000	3.8
Aroclor-1260-2	6.408	6.311	6.511	483.890	500.000	-3.2
Aroclor-1260-3	6.764	6.667	6.867	475.490	500.000	-4.9
Aroclor-1260-4	6.982	6.886	7.086	439.730	500.000	-12.1
Aroclor-1260-5	7.295	7.199	7.399	446.510	500.000	-10.7
Decachlorobiphenyl	8.695	8.599	8.799	46.760	50.000	-6.5
Tetrachloro-m-xylene	3.461	3.363	3.563	50.950	50.000	1.9



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/19/2023 09/19/2023

Client Sample No.: CCAL01 Date Analyzed: 10/09/2023

Lab Sample No.: AR1660CCC500 Data File : PQ063583.D Time Analyzed: 22:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.790	3.693	3.893	476.730	500.000	-4.7
Aroclor-1016-2	3.806	3.709	3.909	487.650	500.000	-2.5
Aroclor-1016-3	3.967	3.870	4.070	490.390	500.000	-1.9
Aroclor-1016-4	4.016	3.920	4.120	496.180	500.000	-0.8
Aroclor-1016-5	4.213	4.116	4.316	493.330	500.000	-1.3
Aroclor-1260-1	5.212	5.116	5.316	480.630	500.000	-3.9
Aroclor-1260-2	5.403	5.306	5.506	476.360	500.000	-4.7
Aroclor-1260-3	5.545	5.450	5.650	462.440	500.000	-7.5
Aroclor-1260-4	6.010	5.914	6.114	467.700	500.000	-6.5
Aroclor-1260-5	6.257	6.160	6.360	467.950	500.000	-6.4
Decachlorobiphenyl	7.580	7.485	7.685	52.650	50.000	5.3
Tetrachloro-m-xylene	2.779	2.681	2.881	50.630	50.000	1.3