



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/19/2023 09/19/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ063594.D Time Analyzed: 01:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.576	4.479	4.679	502.520	500.000	0.5
Aroclor-1016-2	4.595	4.498	4.698	479.510	500.000	-4.1
Aroclor-1016-3	4.653	4.557	4.757	486.610	500.000	-2.7
Aroclor-1016-4	4.746	4.649	4.849	501.620	500.000	0.3
Aroclor-1016-5	5.035	4.939	5.139	494.070	500.000	-1.2
Aroclor-1260-1	6.146	6.049	6.249	468.090	500.000	-6.4
Aroclor-1260-2	6.407	6.311	6.511	448.350	500.000	-10.3
Aroclor-1260-3	6.763	6.667	6.867	480.570	500.000	-3.9
Aroclor-1260-4	6.982	6.886	7.086	438.510	500.000	-12.3
Aroclor-1260-5	7.295	7.199	7.399	442.900	500.000	-11.4
Decachlorobiphenyl	8.694	8.599	8.799	46.740	50.000	-6.5
Tetrachloro-m-xylene	3.460	3.363	3.563	50.690	50.000	1.4



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/19/2023 09/19/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ063594.D Time Analyzed: 01:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.790	3.693	3.893	492.490	500.000	-1.5
Aroclor-1016-2	3.806	3.709	3.909	498.420	500.000	-0.3
Aroclor-1016-3	3.967	3.870	4.070	518.970	500.000	3.8
Aroclor-1016-4	4.016	3.920	4.120	547.740	500.000	9.5
Aroclor-1016-5	4.212	4.116	4.316	517.220	500.000	3.4
Aroclor-1260-1	5.211	5.116	5.316	491.340	500.000	-1.7
Aroclor-1260-2	5.402	5.306	5.506	474.410	500.000	-5.1
Aroclor-1260-3	5.544	5.450	5.650	463.620	500.000	-7.3
Aroclor-1260-4	6.010	5.914	6.114	486.590	500.000	-2.7
Aroclor-1260-5	6.256	6.160	6.360	481.170	500.000	-3.8
Decachlorobiphenyl	7.579	7.485	7.685	53.610	50.000	7.2
Tetrachloro-m-xylene	2.779	2.681	2.881	51.960	50.000	3.9