



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/10/2023 10/10/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PR063895.D Time Analyzed: 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.980	4.880	5.080	505.180	500.000	1.0
Aroclor-1016-2	5.003	4.903	5.103	514.600	500.000	2.9
Aroclor-1016-3	5.068	4.968	5.168	509.930	500.000	2.0
Aroclor-1016-4	5.172	5.071	5.271	516.880	500.000	3.4
Aroclor-1016-5	5.491	5.391	5.591	525.700	500.000	5.1
Aroclor-1260-1	6.716	6.616	6.816	511.160	500.000	2.2
Aroclor-1260-2	7.001	6.902	7.102	511.390	500.000	2.3
Aroclor-1260-3	7.394	7.295	7.495	523.460	500.000	4.7
Aroclor-1260-4	7.639	7.539	7.739	501.300	500.000	0.3
Aroclor-1260-5	7.980	7.880	8.080	518.180	500.000	3.6
Decachlorobiphenyl	9.764	9.665	9.865	50.690	50.000	1.4
Tetrachloro-m-xylene	3.737	3.637	3.837	52.290	50.000	4.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/10/2023 10/10/2023

Client Sample No.: CCAL02 Date Analyzed: 10/10/2023

Lab Sample No.: AR1660CCC500 Data File : PR063895.D Time Analyzed: 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.179	4.080	4.280	507.860	500.000	1.6
Aroclor-1016-2	4.199	4.099	4.299	513.120	500.000	2.6
Aroclor-1016-3	4.382	4.282	4.482	509.790	500.000	2.0
Aroclor-1016-4	4.434	4.335	4.535	507.810	500.000	1.6
Aroclor-1016-5	4.658	4.558	4.758	516.490	500.000	3.3
Aroclor-1260-1	5.778	5.678	5.878	514.870	500.000	3.0
Aroclor-1260-2	5.987	5.887	6.087	504.830	500.000	1.0
Aroclor-1260-3	6.149	6.050	6.250	482.480	500.000	-3.5
Aroclor-1260-4	6.668	6.568	6.768	495.260	500.000	-0.9
Aroclor-1260-5	6.937	6.837	7.037	487.100	500.000	-2.6
Decachlorobiphenyl	8.437	8.337	8.537	51.290	50.000	2.6
Tetrachloro-m-xylene	3.027	2.927	3.127	52.070	50.000	4.1