



Surrogate Summary

SDG No.: O4699Client: Chemtech Consulting GroupAnalytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP060248.D	PIBLK-PP060248.D	Tetrachloro-m-xylene	1	20	19.4	97		60	140
		Decachlorobiphenyl	1	20	21.8	109		60	140
		Tetrachloro-m-xylene	2	20	20.2	101		60	140
		Decachlorobiphenyl	2	20	21.0	105		60	140
I.BLK-PP060802.D	PIBLK-PP060802.D	Tetrachloro-m-xylene	1	20	20.1	101		60	140
		Decachlorobiphenyl	1	20	21.9	110		60	140
		Tetrachloro-m-xylene	2	20	21.1	106		60	140
		Decachlorobiphenyl	2	20	19.7	99		60	140
O4699-08	LOQ-WATER-02-QT4-2023	Tetrachloro-m-xylene	1	20	21.2	106		21	155
		Decachlorobiphenyl	1	20	23.2	116		10	173
		Tetrachloro-m-xylene	2	20	22.1	111		21	155
		Decachlorobiphenyl	2	20	21.5	107		10	173
I.BLK-PP060813.D	PIBLK-PP060813.D	Tetrachloro-m-xylene	1	20	20.1	101		60	140
		Decachlorobiphenyl	1	20	22.1	111		60	140
		Tetrachloro-m-xylene	2	20	21.2	106		60	140
		Decachlorobiphenyl	2	20	20.6	103		60	140
I.BLK-PP060838.D	PIBLK-PP060838.D	Tetrachloro-m-xylene	1	20	21.1	105		60	140
		Decachlorobiphenyl	1	20	22.9	114		60	140
		Tetrachloro-m-xylene	2	20	22.1	111		60	140
		Decachlorobiphenyl	2	20	21.6	108		60	140
O4699-08	LOQ-WATER-02-QT4-2023	Tetrachloro-m-xylene	1	20	20.9	104		21	155
		Decachlorobiphenyl	1	20	23.5	117		10	173
		Tetrachloro-m-xylene	2	20	21.8	109		21	155
		Decachlorobiphenyl	2	20	21.8	109		10	173
I.BLK-PP060852.D	PIBLK-PP060852.D	Tetrachloro-m-xylene	1	20	21.5	108		60	140
		Decachlorobiphenyl	1	20	23.6	118		60	140
		Tetrachloro-m-xylene	2	20	22.5	113		60	140
		Decachlorobiphenyl	2	20	22.3	111		60	140