



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Instrument ID: ECD_L Date(s) Analyzed: 10/03/2023 10/03/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.06	5.96	6.16	17358100
		2	6.46	6.36	6.56	17767900
		3	6.97	6.87	7.07	63718800
		4	7.07	6.97	7.17	37439400
		5	7.76	7.66	7.86	45530400



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Instrument ID: ECD_L Date(s) Analyzed: 10/03/2023 10/03/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.17	5.07	5.27	5595660
		2	5.53	5.43	5.63	6373840
		3	6.44	6.34	6.54	24251700
		4	6.57	6.47	6.67	34769900
		5	6.88	6.78	6.98	22876900