

Report of Analysis

Client: Chemtech Consulting Group

1000

Units:

mL

Date Collected: 10/02/23

Project: LOD-LOQ Study

Date Received: 10/02/23

Client Sample ID: LOD-MDL-WATER-01-QT4-2023

SDG No.: 04699

Lab Sample ID: O4699-07

Matrix: WATER

Analytical Method: SW8081

1000000 uL

Soil Aliquot Vol: uL

Test: Pesticide-TCL

Extraction Type:

Sample Wt/Vol:

Injection Volume:

% Solid:

Final Vol:

GPC Factor: 1.0 PH:

Prep Method: 3510C

File ID/Qc Batch: Dilution:

Prep Date Date Analyzed

Prep Batch ID

Decanted:

PL085801.D 1

10/09/23 08:50

10/09/23 22:07 PB156248

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	2.10	J	0.78	5.00	ug/L
319-85-7	beta-BHC	3.10	J	1.40	5.00	ug/L
319-86-8	delta-BHC	2.40	J	1.20	5.00	ug/L ug/L
58-89-9	gamma-BHC (Lindane)	2.20	J	0.64	5.00	ug/L
76-44-8	Heptachlor	2.60	J	0.73	5.00	ug/L ug/L
309-00-2	Aldrin	2.30	J	0.71	5.00	ug/L
1024-57-3	Heptachlor epoxide	2.60	J	1.00	5.00	ug/L ug/L
959-98-8	Endosulfan I	2.70	J	0.59	5.00	ug/L
60-57-1	Dieldrin	2.50	J	0.54	5.00	ug/L ug/L
72-55-9	4,4-DDE	2.70	J	0.59	5.00	ug/L ug/L
72-20-8	Endrin	2.60	J	0.43	5.00	ug/L
33213-65-9	Endosulfan II	3.20	J	0.76	5.00	ug/L
72-54-8	4,4-DDD	3.10	J	0.84	5.00	ug/L
1031-07-8	Endosulfan Sulfate	3.10	J	0.51	5.00	ug/L ug/L
50-29-3	4,4-DDT	2.80	J	0.65	5.00	ug/L
72-43-5	Methoxychlor	3.90	JР	0.66	5.00	ug/L ug/L
53494-70-5	Endrin ketone	3.00	J	0.84	5.00	ug/L
7421-93-4	Endrin aldehyde	3.40	J	0.70	5.00	ug/L ug/L
5103-71-9	alpha-Chlordane	2.70	J	0.74	5.00	ug/L ug/L
5103-74-2	gamma-Chlordane	2.70	J	0.65	5.00	ug/L ug/L
8001-35-2	Toxaphene	18.0	U	18.0	100	ug/L
0001 33 2	Толириене	10.0	C	10.0	100	ид/ Е
SURROGATES						
2051-24-3	Decachlorobiphenyl	23.9		27 - 142	119%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.9		35 - 148	95%	SPK: 20



Project:

Date Collected:

Date Received:

SDG No.:

Final Vol:

10/09/23 22:07

Injection Volume:

Test:

Matrix: % Solid:

10/02/23

10/02/23

O4699

WATER

1000000

Pesticide-TCL

Decanted:

иL

Prep Batch ID

PB156248

Report of Analysis

Client: Chemtech Consulting Group

LOD-LOQ Study

Client Sample ID: LOD-MDL-WATER-01-QT4-2023

Lab Sample ID: O4699-07

Analytical Method: SW8081

Sample Wt/Vol: 1000 Units: mL

Soil Aliquot Vol: uL

Extraction Type:

PL085801.D

GPC Factor: 1.0 PH:

1

Prep Method: 3510C

File ID/Qc Batch: Dilution: Prep Date Date Analyzed

CAS Number Parameter Conc. Qualifier MDL LOQ / CRQL Units

10/09/23 08:50

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit