



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG No.: O4699

Calibration Sequence : FE100323		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	232786857	136933	FE044569.D
850	119810556	140954	FE044570.D
340	49042674	144243	FE044571.D
170	26900997	158241	FE044572.D
85	13096844	154081	FE044573.D
AVG RF : 146890		% RSD : 6.108	AVG RT : 15.0158