



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG No.: O4699

Calibration Sequence : FG092123		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	169250052	99559	FG012937.D
850	86065373	101253	FG012938.D
340	35935382	105692	FG012939.D
170	19777002	116335	FG012940.D
85	9725826	114421	FG012941.D
AVG RF : 107452		% RSD : 7.077	AVG RT : 15.078