



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG No.: O4699
DataFile: FE044685.D Analyst Name: YP\AJ Analyst Date: 10-07-2023

Conc. (PPM)	Area Count	RF	Average RF	%D
850	116994271	137640	146890	6.297