



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Instrument ID: ECD_R Calibration Date(s): 10/03/2023 10/03/2023

Calibration Times: 11:29 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR063811.D</u>		CF 0.25 = <u>PR063812.D</u>			
CF 0.1 = <u>PR063813.D</u>		CF 0.05 = <u>PR063814.D</u>		CF 0.025 = <u>PR063815.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	158271000000	170704000000	162595000000	159553000000	153308000000	160886000000	4
EDB	720353000000	761559000000	721601000000	721330000000	719500000000	728868000000	3



CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract:

CHEM02

Lab Code:

CHEM

Case No.:

O4699

SAS No.:

O4699

SDG NO.:

O4699

Instrument ID:

ECD_R

Calibration Date(s):

10/03/2023

10/03/2023

Calibration Times:

11:29

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = PR063811.D		CF 0.25 = PR063812.D			
CF 0.1 = PR063813.D		CF 0.05 = PR063814.D		CF 0.025 = PR063815.D			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	95256800000	100121000000	96259600000	90632900000	98159600000	96086000000	4
EDB	43192000000	47003000000	44786800000	46010500000	42490600000	44696600000	4