



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/03/2023 Initial Calibration Date(s): 10/03/2023 10/03/2023

Continuing Calib Time: 14:43 Initial Calibration Time(s): 11:29 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	6.64	6.64	6.54	6.74	0.00
EDB	2.85	2.85	2.75	2.95	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/03/2023 Initial Calibration Date(s): 10/03/2023 10/03/2023

Continuing Calib Time: 14:43 Initial Calibration Time(s): 11:29 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	5.42	5.42	5.32	5.52	0.00
EDB	2.10	2.10	2.00	2.20	0.00