



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/03/2023 10/03/2023

Client Sample No.: CCAL02 Date Analyzed: 10/03/2023

Lab Sample No.: M8011-504CCC 0.1 Data File : PR063827.D Time Analyzed: 14:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.640	6.540	6.740	0.100	0.100	0.0
EDB	2.845	2.745	2.945	0.100	0.100	0.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/03/2023 10/03/2023

Client Sample No.: CCAL02 Date Analyzed: 10/03/2023

Lab Sample No.: M8011-504CCC 0.1 Data File : PR063827.D Time Analyzed: 14:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.421	5.318	5.518	0.100	0.100	0.0
EDB	2.100	1.999	2.199	0.100	0.100	0.0