



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Initial and Continuing Calibration Verification

Client: RMJ Environomics, Inc.

SDG No.: O5252

Project: 245 Greenwood Ave

RunNo.: LB128256

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.096	0.099	97	90-110	11/09/2023
Sample ID:	CCV1						
Cyanide		mg/L	0.25	0.25	100	90-110	11/09/2023
Sample ID:	CCV2						
Cyanide		mg/L	0.25	0.25	100	90-110	11/09/2023
Sample ID:	CCV3						
Cyanide		mg/L	0.25	0.25	100	90-110	11/09/2023
Sample ID:	CCV4						
Cyanide		mg/L	0.26	0.25	104	90-110	11/09/2023