



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: RMJ Environomics, Inc.

SDG No.: O5252

Contract: RMJE02

Lab Code: CHEM

Case No.: O5252

SAS No.: O5252

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	83	40 - 160	CV	11/07/2023	16:21	LB128217
CRI01	Aluminum	96.1	100	96	40 - 160	P	11/09/2023	11:28	LB128258
	Antimony	50.2	50.0	100	40 - 160	P	11/09/2023	11:28	LB128258
	Arsenic	23.5	20.0	117	40 - 160	P	11/09/2023	11:28	LB128258
	Barium	91.2	100	91	40 - 160	P	11/09/2023	11:28	LB128258
	Beryllium	5.95	6.0	99	40 - 160	P	11/09/2023	11:28	LB128258
	Cadmium	6.35	6.0	106	40 - 160	P	11/09/2023	11:28	LB128258
	Calcium	1970	2000	99	40 - 160	P	11/09/2023	11:28	LB128258
	Chromium	9.44	10.0	94	40 - 160	P	11/09/2023	11:28	LB128258
	Cobalt	28.9	30.0	96	40 - 160	P	11/09/2023	11:28	LB128258
	Copper	21.3	20.0	106	40 - 160	P	11/09/2023	11:28	LB128258
	Iron	97.1	100	97	40 - 160	P	11/09/2023	11:28	LB128258
	Lead	11.9	12.0	99	40 - 160	P	11/09/2023	11:28	LB128258
	Magnesium	2110	2000	106	40 - 160	P	11/09/2023	11:28	LB128258
	Manganese	19.8	20.0	99	40 - 160	P	11/09/2023	11:28	LB128258
	Nickel	39.6	40.0	99	40 - 160	P	11/09/2023	11:28	LB128258
	Potassium	2000	2000	100	40 - 160	P	11/09/2023	11:28	LB128258
	Selenium	21.8	20.0	109	40 - 160	P	11/09/2023	11:28	LB128258
	Silver	9.54	10.0	95	40 - 160	P	11/09/2023	11:28	LB128258
	Sodium	2000	2000	100	40 - 160	P	11/09/2023	11:28	LB128258
	Thallium	30.6	40.0	77	40 - 160	P	11/09/2023	11:28	LB128258
	Vanadium	38.9	40.0	97	40 - 160	P	11/09/2023	11:28	LB128258
	Zinc	40.7	40.0	102	40 - 160	P	11/09/2023	11:28	LB128258