



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC110723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC065037.D

Aliphatic C9-C12	11419136.000	60.000	3.211	6.483	190318.933	190794.450	0.249
Aliphatic C12-C16	7505684.000	40.000	6.484	9.865	187642.100	188856.476	0.643
Aliphatic C16-C21	10583830.000	60.000	9.866	13.214	176397.167	181359.642	2.736
Aliphatic C21-C28	12712462.000	80.000	13.215	16.858	158905.775	163834.452	3.008
Aliphatic C28-C40	18660728.000	120.000	16.859	21.689	155506.067	152616.945	-1.893
Aliphatic EPH	60881840.000	360.000	3.211	21.689	169116.222	170289.706	0.689