



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC110723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC065057.D

Aliphatic C9-C12	11973745.000	60.000	3.211	6.483	199562.417	190794.450	-4.596
Aliphatic C12-C16	7732592.000	40.000	6.484	9.865	193314.800	188856.476	-2.361
Aliphatic C16-C21	10770619.000	60.000	9.866	13.214	179510.317	181359.642	1.020
Aliphatic C21-C28	12900481.000	80.000	13.215	16.858	161256.013	163834.452	1.574
Aliphatic C28-C40	18438248.000	120.000	16.859	21.689	153652.067	152616.945	-0.678
Aliphatic EPH	61815685.000	360.000	3.211	21.689	171710.236	170289.706	-0.834