



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC110823AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC065060.D

Aliphatic C9-C12	12192589.000	60.000	3.210	6.479	203209.817	190794.450	-6.507
Aliphatic C12-C16	7855125.000	40.000	6.480	9.862	196378.125	188856.476	-3.983
Aliphatic C16-C21	10932343.000	60.000	9.863	13.209	182205.717	181359.642	-0.467
Aliphatic C21-C28	13078488.000	80.000	13.210	16.852	163481.100	163834.452	0.216
Aliphatic C28-C40	18666955.000	120.000	16.853	21.679	155557.958	152616.945	-1.927
Aliphatic EPH	62725500.000	360.000	3.210	21.679	174237.500	170289.706	-2.318