



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC110823AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC065064.D

Aliphatic C9-C12	12004076.000	60.000	3.210	6.479	200067.933	190794.450	-4.860
Aliphatic C12-C16	7833771.000	40.000	6.480	9.862	195844.275	188856.476	-3.700
Aliphatic C16-C21	10957602.000	60.000	9.863	13.209	182626.700	181359.642	-0.699
Aliphatic C21-C28	13143703.000	80.000	13.210	16.852	164296.288	163834.452	-0.282
Aliphatic C28-C40	18759699.000	120.000	16.853	21.679	156330.825	152616.945	-2.433
Aliphatic EPH	62698851.000	360.000	3.210	21.679	174163.475	170289.706	-2.275