



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD110723AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046838.D

Aromatic C10-C12	7666603.000	40.000	4.196	5.923	191665.075	196684.329	2.552
Aromatic C12-C16	11384136.000	60.000	5.924	8.535	189735.600	193599.401	1.996
Aromatic C16-C21	13373136.000	80.000	8.536	12.798	167164.200	174891.939	4.419
Aromatic C21-C36	24866167.000	180.000	12.799	18.185	138145.372	144967.841	4.706
Aromatic EPH	57290042.000	360.000	4.196	18.185	159139.006	165469.177	3.826