



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD110723AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046858.D

Aromatic C10-C12	7968125.000	40.000	4.196	5.923	199203.125	196684.329	-1.281
Aromatic C12-C16	11850876.000	60.000	5.924	8.535	197514.600	193599.401	-2.022
Aromatic C16-C21	14143623.000	80.000	8.536	12.798	176795.288	174891.939	-1.088
Aromatic C21-C36	26730883.000	180.000	12.799	18.185	148504.906	144967.841	-2.440
Aromatic EPH	60693507.000	360.000	4.196	18.185	168593.075	165469.177	-1.888