



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/07/2023 Initial Calibration Date(s): 10/24/2023 10/25/2023

Continuing Calib Time: 09:20 Initial Calibration Time(s): 21:19 04:56

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1016-2	(2)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.73	5.74	5.64	5.84	0.01
Aroclor-1016-4	(4)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-5	(5)	6.12	6.13	6.03	6.23	0.01
Aroclor-1260-1	(1)	7.25	7.26	7.16	7.36	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.10	8.10	8.00	8.20	0.00
Aroclor-1260-5	(5)	8.42	8.43	8.33	8.53	0.01
Tetrachloro-m-xylene		4.47	4.48	4.38	4.58	0.01
Decachlorobiphenyl		10.27	10.28	10.18	10.38	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/07/2023 Initial Calibration Date(s): 10/24/2023 10/25/2023

Continuing Calib Time: 09:20 Initial Calibration Time(s): 21:19 04:56

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.70	4.71	4.61	4.81	0.01
Aroclor-1016-2	(2)	4.72	4.73	4.63	4.83	0.01
Aroclor-1016-3	(3)	4.90	4.91	4.81	5.01	0.01
Aroclor-1016-4	(4)	4.94	4.95	4.85	5.05	0.01
Aroclor-1016-5	(5)	5.15	5.16	5.06	5.26	0.01
Aroclor-1260-1	(1)	6.18	6.20	6.10	6.30	0.02
Aroclor-1260-2	(2)	6.37	6.38	6.28	6.48	0.01
Aroclor-1260-3	(3)	6.52	6.54	6.44	6.64	0.02
Aroclor-1260-4	(4)	7.00	7.01	6.91	7.11	0.01
Aroclor-1260-5	(5)	7.24	7.25	7.15	7.35	0.01
Tetrachloro-m-xylene		3.62	3.63	3.53	3.73	0.01
Decachlorobiphenyl		8.62	8.63	8.53	8.73	0.02