



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 10:05 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1016-2	(2)	5.78	5.78	5.68	5.88	0.00
Aroclor-1016-3	(3)	5.84	5.84	5.74	5.94	0.00
Aroclor-1016-4	(4)	5.94	5.94	5.84	6.04	0.00
Aroclor-1016-5	(5)	6.24	6.24	6.14	6.34	0.00
Aroclor-1260-1	(1)	7.39	7.38	7.28	7.48	-0.01
Aroclor-1260-2	(2)	7.65	7.64	7.54	7.74	-0.01
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.58	8.58	8.48	8.68	0.00
Tetrachloro-m-xylene		4.57	4.56	4.46	4.66	-0.01
Decachlorobiphenyl		10.55	10.54	10.44	10.64	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 10:05 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.80	4.79	4.69	4.89	-0.01
Aroclor-1016-2	(2)	4.82	4.81	4.71	4.91	-0.01
Aroclor-1016-3	(3)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-4	(4)	5.04	5.03	4.93	5.13	-0.01
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.30	6.29	6.19	6.39	-0.01
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.64	6.64	6.54	6.74	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.36	7.35	7.25	7.45	-0.01
Tetrachloro-m-xylene		3.70	3.69	3.59	3.79	-0.01
Decachlorobiphenyl		8.77	8.76	8.66	8.86	-0.01