



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 20:21 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.77	5.75	5.65	5.85	-0.02
Aroclor-1016-2	(2)	5.79	5.78	5.68	5.88	-0.01
Aroclor-1016-3	(3)	5.86	5.84	5.74	5.94	-0.02
Aroclor-1016-4	(4)	5.96	5.94	5.84	6.04	-0.02
Aroclor-1016-5	(5)	6.26	6.24	6.14	6.34	-0.02
Aroclor-1260-1	(1)	7.40	7.38	7.28	7.48	-0.02
Aroclor-1260-2	(2)	7.66	7.64	7.54	7.74	-0.02
Aroclor-1260-3	(3)	8.03	8.01	7.91	8.11	-0.01
Aroclor-1260-4	(4)	8.26	8.24	8.14	8.34	-0.02
Aroclor-1260-5	(5)	8.60	8.58	8.48	8.68	-0.02
Tetrachloro-m-xylene		4.58	4.56	4.46	4.66	-0.02
Decachlorobiphenyl		10.57	10.54	10.44	10.64	-0.03



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 20:21 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.81	4.79	4.69	4.89	-0.02
Aroclor-1016-2 (2)	4.83	4.81	4.71	4.91	-0.02
Aroclor-1016-3 (3)	5.01	4.99	4.89	5.09	-0.02
Aroclor-1016-4 (4)	5.05	5.03	4.93	5.13	-0.02
Aroclor-1016-5 (5)	5.27	5.25	5.15	5.35	-0.02
Aroclor-1260-1 (1)	6.31	6.29	6.19	6.39	-0.02
Aroclor-1260-2 (2)	6.50	6.48	6.38	6.58	-0.02
Aroclor-1260-3 (3)	6.65	6.64	6.54	6.74	-0.01
Aroclor-1260-4 (4)	7.13	7.11	7.01	7.21	-0.02
Aroclor-1260-5 (5)	7.37	7.35	7.25	7.45	-0.02
Tetrachloro-m-xylene	3.71	3.69	3.59	3.79	-0.02
Decachlorobiphenyl	8.79	8.76	8.66	8.86	-0.03