



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 20:37 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.77	5.76	5.66	5.86	-0.01
Aroclor-1242-2 (2)	5.79	5.78	5.68	5.88	-0.01
Aroclor-1242-3 (3)	5.86	5.84	5.74	5.94	-0.02
Aroclor-1242-4 (4)	5.96	5.94	5.84	6.04	-0.02
Aroclor-1242-5 (5)	6.71	6.69	6.59	6.79	-0.02
Tetrachloro-m-xylene	4.58	4.56	4.46	4.66	-0.02
Decachlorobiphenyl	10.58	10.54	10.44	10.64	-0.04



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 20:37 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.81	4.80	4.70	4.90	-0.01
Aroclor-1242-2 (2)	4.83	4.82	4.72	4.92	-0.01
Aroclor-1242-3 (3)	5.01	4.99	4.89	5.09	-0.02
Aroclor-1242-4 (4)	5.09	5.08	4.98	5.18	-0.01
Aroclor-1242-5 (5)	5.62	5.61	5.51	5.71	-0.01
Tetrachloro-m-xylene	3.71	3.70	3.60	3.80	-0.01
Decachlorobiphenyl	8.79	8.77	8.67	8.87	-0.01