



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 20:54 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.77	5.75	5.65	5.85	-0.02
Aroclor-1248-2 (2)	6.05	6.03	5.93	6.13	-0.02
Aroclor-1248-3 (3)	6.26	6.24	6.14	6.34	-0.02
Aroclor-1248-4 (4)	6.67	6.65	6.55	6.75	-0.02
Aroclor-1248-5 (5)	6.71	6.69	6.59	6.79	-0.02
Tetrachloro-m-xylene	4.58	4.56	4.46	4.66	-0.01
Decachlorobiphenyl	10.57	10.54	10.44	10.64	-0.03



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 20:54 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.81	4.79	4.69	4.89	-0.02
Aroclor-1248-2 (2)	5.05	5.03	4.93	5.13	-0.02
Aroclor-1248-3 (3)	5.09	5.07	4.97	5.17	-0.02
Aroclor-1248-4 (4)	5.26	5.25	5.15	5.35	-0.01
Aroclor-1248-5 (5)	5.66	5.64	5.54	5.74	-0.02
Tetrachloro-m-xylene	3.71	3.69	3.59	3.79	-0.02
Decachlorobiphenyl	8.78	8.76	8.66	8.86	-0.02