



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 21:10 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.64	6.62	6.52	6.72	-0.02
Aroclor-1254-2 (2)	6.86	6.84	6.74	6.94	-0.02
Aroclor-1254-3 (3)	7.23	7.21	7.11	7.31	-0.02
Aroclor-1254-4 (4)	7.52	7.50	7.40	7.60	-0.02
Aroclor-1254-5 (5)	7.94	7.93	7.83	8.03	-0.01
Tetrachloro-m-xylene	4.57	4.56	4.46	4.66	-0.01
Decachlorobiphenyl	10.57	10.53	10.43	10.63	-0.04



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Continuing Calib Date: 11/06/2023 Initial Calibration Date(s): 10/27/2023 10/27/2023

Continuing Calib Time: 21:10 Initial Calibration Time(s): 11:03 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1	(1)	5.62	5.60	5.50	5.70	-0.02
Aroclor-1254-2	(2)	5.77	5.75	5.65	5.85	-0.02
Aroclor-1254-3	(3)	6.18	6.16	6.06	6.26	-0.02
Aroclor-1254-4	(4)	6.41	6.39	6.29	6.49	-0.02
Aroclor-1254-5	(5)	6.83	6.81	6.71	6.91	-0.02
Tetrachloro-m-xylene		3.71	3.69	3.59	3.79	-0.02
Decachlorobiphenyl		8.78	8.76	8.66	8.86	-0.02