



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/24/2023 10/24/2023

Client Sample No.: CCAL01 Date Analyzed: 11/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO099451.D Time Analyzed: 09:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.643	5.550	5.750	457.780	500.000	-8.4
Aroclor-1016-2	5.665	5.574	5.774	455.560	500.000	-8.9
Aroclor-1016-3	5.728	5.635	5.835	462.180	500.000	-7.6
Aroclor-1016-4	5.826	5.734	5.934	466.240	500.000	-6.8
Aroclor-1016-5	6.122	6.030	6.230	444.680	500.000	-11.1
Aroclor-1260-1	7.252	7.160	7.360	448.120	500.000	-10.4
Aroclor-1260-2	7.510	7.417	7.617	467.890	500.000	-6.4
Aroclor-1260-3	7.870	7.777	7.977	466.020	500.000	-6.8
Aroclor-1260-4	8.097	8.004	8.204	472.330	500.000	-5.5
Aroclor-1260-5	8.421	8.329	8.529	501.630	500.000	0.3
Decachlorobiphenyl	10.268	10.181	10.381	51.480	50.000	3.0
Tetrachloro-m-xylene	4.469	4.377	4.577	43.000	50.000	-14.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/24/2023 10/24/2023

Client Sample No.: CCAL01 Date Analyzed: 11/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO099451.D Time Analyzed: 09:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.701	4.613	4.813	459.150	500.000	-8.2
Aroclor-1016-2	4.720	4.631	4.831	468.650	500.000	-6.3
Aroclor-1016-3	4.896	4.807	5.007	460.170	500.000	-8.0
Aroclor-1016-4	4.938	4.849	5.049	442.230	500.000	-11.6
Aroclor-1016-5	5.151	5.062	5.262	446.690	500.000	-10.7
Aroclor-1260-1	6.183	6.095	6.295	445.860	500.000	-10.8
Aroclor-1260-2	6.372	6.283	6.483	459.510	500.000	-8.1
Aroclor-1260-3	6.524	6.436	6.636	453.910	500.000	-9.2
Aroclor-1260-4	6.996	6.907	7.107	450.830	500.000	-9.8
Aroclor-1260-5	7.238	7.150	7.350	464.260	500.000	-7.1
Decachlorobiphenyl	8.615	8.530	8.730	42.970	50.000	-14.1
Tetrachloro-m-xylene	3.622	3.531	3.731	44.680	50.000	-10.6