



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/24/2023 10/24/2023

Client Sample No.: CCAL02 Date Analyzed: 11/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO099463.D Time Analyzed: 13:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.642	5.550	5.750	452.020	500.000	-9.6
Aroclor-1016-2	5.665	5.574	5.774	434.890	500.000	-13.0
Aroclor-1016-3	5.726	5.635	5.835	444.140	500.000	-11.2
Aroclor-1016-4	5.824	5.734	5.934	444.900	500.000	-11.0
Aroclor-1016-5	6.121	6.030	6.230	428.830	500.000	-14.2
Aroclor-1260-1	7.251	7.160	7.360	426.000	500.000	-14.8
Aroclor-1260-2	7.508	7.417	7.617	443.460	500.000	-11.3
Aroclor-1260-3	7.868	7.777	7.977	430.760	500.000	-13.8
Aroclor-1260-4	8.094	8.004	8.204	450.950	500.000	-9.8
Aroclor-1260-5	8.419	8.329	8.529	483.230	500.000	-3.4
Decachlorobiphenyl	10.267	10.181	10.381	48.820	50.000	-2.4
Tetrachloro-m-xylene	4.469	4.377	4.577	42.520	50.000	-15.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/24/2023 10/24/2023

Client Sample No.: CCAL02 Date Analyzed: 11/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO099463.D Time Analyzed: 13:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.703	4.613	4.813	474.980	500.000	-5.0
Aroclor-1016-2	4.722	4.631	4.831	494.090	500.000	-1.2
Aroclor-1016-3	4.897	4.807	5.007	485.470	500.000	-2.9
Aroclor-1016-4	4.939	4.849	5.049	457.010	500.000	-8.6
Aroclor-1016-5	5.152	5.062	5.262	467.430	500.000	-6.5
Aroclor-1260-1	6.184	6.095	6.295	441.960	500.000	-11.6
Aroclor-1260-2	6.372	6.283	6.483	458.070	500.000	-8.4
Aroclor-1260-3	6.525	6.436	6.636	443.170	500.000	-11.4
Aroclor-1260-4	6.996	6.907	7.107	435.620	500.000	-12.9
Aroclor-1260-5	7.239	7.150	7.350	454.500	500.000	-9.1
Decachlorobiphenyl	8.616	8.530	8.730	41.900	50.000	-16.2
Tetrachloro-m-xylene	3.624	3.531	3.731	45.450	50.000	-9.1