



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/27/2023 10/27/2023

Client Sample No.: CCAL03 Date Analyzed: 11/06/2023

Lab Sample No.: AR1660CCC500 Data File : PP061440.D Time Analyzed: 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.758	5.652	5.852	533.060	500.000	6.6
Aroclor-1016-2	5.781	5.676	5.876	533.840	500.000	6.8
Aroclor-1016-3	5.844	5.739	5.939	529.500	500.000	5.9
Aroclor-1016-4	5.944	5.839	6.039	532.720	500.000	6.5
Aroclor-1016-5	6.244	6.139	6.339	547.570	500.000	9.5
Aroclor-1260-1	7.386	7.282	7.482	449.250	500.000	-10.2
Aroclor-1260-2	7.646	7.541	7.741	464.630	500.000	-7.1
Aroclor-1260-3	8.011	7.905	8.105	487.930	500.000	-2.4
Aroclor-1260-4	8.244	8.139	8.339	493.070	500.000	-1.4
Aroclor-1260-5	8.584	8.477	8.677	487.980	500.000	-2.4
Decachlorobiphenyl	10.547	10.436	10.636	46.450	50.000	-7.1
Tetrachloro-m-xylene	4.565	4.459	4.659	55.170	50.000	10.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/27/2023 10/27/2023

Client Sample No.: CCAL03 Date Analyzed: 11/06/2023

Lab Sample No.: AR1660CCC500 Data File : PP061440.D Time Analyzed: 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.797	4.694	4.894	515.070	500.000	3.0
Aroclor-1016-2	4.816	4.712	4.912	523.530	500.000	4.7
Aroclor-1016-3	4.994	4.890	5.090	524.040	500.000	4.8
Aroclor-1016-4	5.037	4.933	5.133	522.060	500.000	4.4
Aroclor-1016-5	5.252	5.148	5.348	511.460	500.000	2.3
Aroclor-1260-1	6.295	6.191	6.391	498.100	500.000	-0.4
Aroclor-1260-2	6.484	6.380	6.580	489.450	500.000	-2.1
Aroclor-1260-3	6.639	6.535	6.735	498.710	500.000	-0.3
Aroclor-1260-4	7.114	7.011	7.211	473.150	500.000	-5.4
Aroclor-1260-5	7.357	7.253	7.453	469.160	500.000	-6.2
Decachlorobiphenyl	8.769	8.664	8.864	42.330	50.000	-15.3
Tetrachloro-m-xylene	3.698	3.594	3.794	53.660	50.000	7.3