



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/27/2023 10/27/2023

Client Sample No.: CCAL04 Date Analyzed: 11/06/2023

Lab Sample No.: AR1660CCC500 Data File : PP061455.D Time Analyzed: 16:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.763	5.652	5.852	523.540	500.000	4.7
Aroclor-1016-2	5.786	5.676	5.876	519.710	500.000	3.9
Aroclor-1016-3	5.849	5.739	5.939	511.240	500.000	2.2
Aroclor-1016-4	5.950	5.839	6.039	526.170	500.000	5.2
Aroclor-1016-5	6.250	6.139	6.339	533.090	500.000	6.6
Aroclor-1260-1	7.393	7.282	7.482	463.530	500.000	-7.3
Aroclor-1260-2	7.652	7.541	7.741	447.410	500.000	-10.5
Aroclor-1260-3	8.016	7.905	8.105	426.220	500.000	-14.8
Aroclor-1260-4	8.251	8.139	8.339	437.170	500.000	-12.6
Aroclor-1260-5	8.590	8.477	8.677	453.000	500.000	-9.4
Decachlorobiphenyl	10.560	10.436	10.636	44.990	50.000	-10.0
Tetrachloro-m-xylene	4.570	4.459	4.659	53.280	50.000	6.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/27/2023 10/27/2023

Client Sample No.: CCAL04 Date Analyzed: 11/06/2023

Lab Sample No.: AR1660CCC500 Data File : PP061455.D Time Analyzed: 16:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.803	4.694	4.894	520.420	500.000	4.1
Aroclor-1016-2	4.822	4.712	4.912	517.580	500.000	3.5
Aroclor-1016-3	5.000	4.890	5.090	524.980	500.000	5.0
Aroclor-1016-4	5.043	4.933	5.133	518.120	500.000	3.6
Aroclor-1016-5	5.258	5.148	5.348	507.640	500.000	1.5
Aroclor-1260-1	6.301	6.191	6.391	484.470	500.000	-3.1
Aroclor-1260-2	6.491	6.380	6.580	471.760	500.000	-5.6
Aroclor-1260-3	6.646	6.535	6.735	482.340	500.000	-3.5
Aroclor-1260-4	7.121	7.011	7.211	466.430	500.000	-6.7
Aroclor-1260-5	7.364	7.253	7.453	456.170	500.000	-8.8
Decachlorobiphenyl	8.778	8.664	8.864	40.720	50.000	-18.6
Tetrachloro-m-xylene	3.703	3.594	3.794	53.730	50.000	7.5