



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Instrument ID: ECD_L Date(s) Analyzed: 11/01/2023 11/01/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.05	5.95	6.15	24773700
		2	6.46	6.36	6.56	20193000
		3	6.97	6.87	7.07	65365100
		4	7.06	6.96	7.16	39592500
		5	7.75	7.65	7.85	46260500



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

Instrument ID: ECD_L Date(s) Analyzed: 11/01/2023 11/01/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.17	5.07	5.27	5774410
		2	5.52	5.42	5.62	6851300
		3	6.44	6.34	6.54	26432000
		4	6.57	6.47	6.67	36861600
		5	6.88	6.78	6.98	25089800