

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/01/2023 11/01/2023

Client Sample No.: CCAL03 Date Analyzed: 11/06/2023

Lab Sample No.: PSTDCCC050 Data File : PL086495.D Time Analyzed: 21:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.535	6.435	6.635	55.970	50.000	11.9
4,4'-DDE	6.015	5.916	6.116	53.360	50.000	6.7
4,4'-DDT	6.849	6.750	6.950	51.640	50.000	3.3
Aldrin	5.064	4.965	5.165	49.950	50.000	-0.1
alpha-BHC	3.806	3.707	3.907	50.180	50.000	0.4
alpha-Chlordane	5.831	5.732	5.932	48.700	50.000	-2.6
beta-BHC	4.341	4.242	4.442	47.970	50.000	-4.1
Decachlorobiphenyl	8.853	8.755	8.955	49.420	50.000	-1.2
delta-BHC	4.584	4.485	4.685	50.450	50.000	0.9
Dieldrin	6.155	6.056	6.256	48.920	50.000	-2.2
Endosulfan I	5.878	5.779	5.979	48.600	50.000	-2.8
Endosulfan II	6.606	6.508	6.708	49.760	50.000	-0.5
Endosulfan sulfate	6.974	6.875	7.075	48.710	50.000	-2.6
Endrin	6.384	6.285	6.485	46.270	50.000	-7.5
Endrin aldehyde	6.738	6.639	6.839	50.040	50.000	0.1
Endrin ketone	7.456	7.357	7.557	49.880	50.000	-0.2
gamma-BHC (Lindane)	4.138	4.039	4.239	50.120	50.000	0.2
gamma-Chlordane	5.751	5.653	5.853	49.160	50.000	-1.7
Heptachlor	4.725	4.626	4.826	50.660	50.000	1.3
Heptachlor epoxide	5.495	5.395	5.595	49.080	50.000	-1.8
Methoxychlor	7.335	7.236	7.436	51.700	50.000	3.4
Tetrachloro-m-xylene	3.353	3.253	3.453	49.860	50.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: RMJE02Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/01/2023 11/01/2023Client Sample No.: CCAL03 Date Analyzed: 11/06/2023Lab Sample No.: PSTDCCC050 Data File : PL086495.D Time Analyzed: 21:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.631	5.532	5.732	57.470	50.000	14.9
4,4'-DDE	5.075	4.976	5.176	55.750	50.000	11.5
4,4'-DDT	5.881	5.783	5.983	53.700	50.000	7.4
Aldrin	4.056	3.957	4.157	52.050	50.000	4.1
alpha-BHC	3.116	3.017	3.217	52.400	50.000	4.8
alpha-Chlordane	4.875	4.777	4.977	49.920	50.000	-0.2
beta-BHC	3.745	3.646	3.846	51.500	50.000	3.0
Decachlorobiphenyl	7.769	7.670	7.870	50.490	50.000	1.0
delta-BHC	3.971	3.872	4.072	52.710	50.000	5.4
Dieldrin	5.194	5.095	5.295	51.560	50.000	3.1
Endosulfan I	4.928	4.829	5.029	48.360	50.000	-3.3
Endosulfan II	5.766	5.666	5.866	51.740	50.000	3.5
Endosulfan sulfate	6.171	6.072	6.272	52.360	50.000	4.7
Endrin	5.467	5.368	5.568	48.970	50.000	-2.1
Endrin aldehyde	5.947	5.848	6.048	53.840	50.000	7.7
Endrin ketone	6.673	6.574	6.774	53.220	50.000	6.4
gamma-BHC (Lindane)	3.443	3.344	3.544	51.600	50.000	3.2
gamma-Chlordane	4.812	4.713	4.913	51.780	50.000	3.6
Heptachlor	3.779	3.680	3.880	51.440	50.000	2.9
Heptachlor epoxide	4.560	4.462	4.662	51.120	50.000	2.2
Methoxychlor	6.466	6.367	6.567	52.740	50.000	5.5
Tetrachloro-m-xylene	2.619	2.518	2.718	51.070	50.000	2.1