

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110623\
 Data File : PL086485.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2023 18:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 20:44:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110123.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 02 05:32:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.353	2.618	154.4E6	48825258	48.662	51.288
28)	SA Decachlor...	8.855	7.768	131.1E6	57619834	49.555	50.233
Target Compounds							
2)	A alpha-BHC	3.807	3.116	232.1E6	68213405	49.172	52.563
3)	MA gamma-BHC...	4.139	3.443	225.9E6	67788418	49.578	51.607
4)	MA Heptachlor	4.726	3.779	215.5E6	66927401	49.436	51.590
5)	MB Aldrin	5.066	4.056	215.0E6	66588599	48.555	52.253
6)	B beta-BHC	4.342	3.746	92587276	29747122	46.942	51.831
7)	B delta-BHC	4.585	3.970	215.0E6	66942197	49.287	52.959
8)	B Heptachlo...	5.496	4.561	185.8E6	61365736	48.039	51.329
9)	A Endosulfan I	5.880	4.928	175.3E6	57812808	47.682	48.713
10)	B gamma-Chl...	5.753	4.812	191.0E6	61387597	48.144	51.875
11)	B alpha-Chl...	5.832	4.876	189.9E6	62697143	47.700	50.189
12)	B 4,4'-DDE	6.016	5.076	164.0E6	54955014	51.675	55.688
13)	MA Dieldrin	6.157	5.194	183.3E6	61818938	47.673	51.767
14)	MA Endrin	6.385	5.468	157.5E6	54738781	47.188	49.675
15)	B Endosulfa...	6.608	5.766	158.2E6	56036159	49.109	51.986
16)	A 4,4'-DDD	6.535	5.632	128.3E6	46568369	54.095	57.176
17)	MA 4,4'-DDT	6.851	5.883	134.5E6	51412166	51.550	54.535
18)	B Endrin al...	6.739	5.948	120.3E6	45028396	49.432	53.395
19)	B Endosulfa...	6.975	6.172	149.6E6	55240883	48.436	52.397
20)	A Methoxychlor	7.336	6.466	73111972	31328267	52.712	53.614
21)	B Endrin ke...	7.457	6.673	158.0E6	65565281	49.516	52.571
22)	Mirex	7.930	6.850	126.4E6	54762423	46.323	47.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL110623\
Data File : PL086485.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2023 18:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 20:44:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL110123.M
Quant Title : GC Extractables
QLast Update : Thu Nov 02 05:32:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

