

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF103023

Review By	yogesh	Review On	10/31/2023 4:29:34 AM				
Supervise By	mohammad	Supervise On	10/31/2023 4:54:30 AM				
SubDirectory	BF103023	HP Acquire Method	BNA_F	HP Processing Method	BF103023		
STD. NAME		STD REF. #					
Tune/Reschk		SP6271					
Initial Calibration Stds		SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236					
CCC		SP6232					
Internal Standard/PEM		S11507 10ul/1000ul sample					
ICV/I.BLK		SP6324					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF136022.D	30 Oct 2023 11:02		CG\JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF136023.D	30 Oct 2023 12:02		CG\JU	Ok
3	SSTDICC005	SSTDICC005	BF136024.D	30 Oct 2023 12:32	Compound#32,54,65 removed from 5 ppm	CG\JU	Ok,M
4	SSTDICC010	SSTDICC010	BF136025.D	30 Oct 2023 13:02	The CALibration is Good for Non-DOD and for 625.1.	CG\JU	Ok
5	SSTDICC020	SSTDICC020	BF136026.D	30 Oct 2023 13:33	Method is Good for DOD (Except Benzidine) and Good for 625.1.	CG\JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF136027.D	30 Oct 2023 14:04	Compound #32,54,65,92 Kept on LR	CG\JU	Ok,M
7	SSTDICC050	SSTDICC050	BF136028.D	30 Oct 2023 14:49		CG\JU	Ok,M
8	SSTDICC060	SSTDICC060	BF136029.D	30 Oct 2023 15:20		CG\JU	Ok,M
9	SSTDICC080	SSTDICC080	BF136030.D	30 Oct 2023 15:51		CG\JU	Ok,M
10	SSTDICV040	ICVBF103023	BF136031.D	30 Oct 2023 16:24		CG\JU	Ok,M
11	PB156520BL	PB156520BL	BF136032.D	30 Oct 2023 17:26		CG\JU	Ok
12	PB156392BL	PB156392BL	BF136033.D	30 Oct 2023 17:56		CG\JU	Ok
13	PB156520BS	PB156520BS	BF136034.D	30 Oct 2023 18:27		CG\JU	Ok,M
14	O5027-01	214	BF136035.D	30 Oct 2023 19:01	Internal standard not added	CG\JU	Not Ok
15	O5090-01	WC-1	BF136036.D	30 Oct 2023 19:32		CG\JU	Ok,M
16	O5062-01	NB-07-102523	BF136037.D	30 Oct 2023 20:02		CG\JU	Ok,M
17	O5073-01	OR-02-102523	BF136038.D	30 Oct 2023 20:33	Need Further 5X	CG\JU	Dilution



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF103023

Review By	yogesh	Review On	10/31/2023 4:29:34 AM					
Supervise By	mohammad	Supervise On	10/31/2023 4:54:30 AM					
SubDirectory	BF103023	HP Acquire Method	BNA_F	HP Processing Method	BF103023			
STD. NAME		STD REF.#						
Tune/Reschk		SP6271						
Initial Calibration Stds		SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236						
CCC		SP6232						
Internal Standard/PEM		S11507 10ul/1000ul sample						
ICV/I.BLK		SP6324						
Surrogate Standard								
MS/MSD Standard								
LCS Standard								

18	O5078-01	EO-03-102323	BF136039.D	30 Oct 2023 21:03		CG\JU	Ok,M
19	O4907-01RE	OR-03-101623RE	BF136040.D	30 Oct 2023 21:33	Fax is already Given	CG\JU	Confirms
20	O4704-02DL	R-1(10-15)DL	BF136041.D	30 Oct 2023 22:04		CG\JU	Ok,M
21	O4704-04DL	R-3(10-15)DL	BF136042.D	30 Oct 2023 22:34		CG\JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110623

Review By	yogesh	Review On	11/6/2023 4:02:34 PM				
Supervise By	mohammad	Supervise On	11/8/2023 1:02:43 AM				
SubDirectory	BF110623	HP Acquire Method	BNA_F	HP Processing Method	BF103023		
STD. NAME		STD REF.#					
Tune/Reschk		SP6271					
Initial Calibration Stds		SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236					
CCC		SP6232					
Internal Standard/PEM		S11510 10ul/1000ul sample					
ICV/I.BLK		SP6324					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF136148.D	06 Nov 2023 11:31		CG\JU	Ok
2	SSTDCCC040	SSTDCCC040	BF136149.D	06 Nov 2023 12:01		CG\JU	Ok,M
3	PB156895BL	PB156895BL	BF136150.D	06 Nov 2023 12:31		CG\JU	Ok
4	O5237-01	111TH-ST-1	BF136151.D	06 Nov 2023 13:26		CG\JU	Ok,M
5	O5243-03	SB-11-AT-4.5-5.0	BF136152.D	06 Nov 2023 13:56		CG\JU	Ok
6	O5234-01	SP-A	BF136153.D	06 Nov 2023 14:26		CG\JU	Ok
7	O5257-09	WC-3	BF136154.D	06 Nov 2023 14:56		CG\JU	Ok
8	O5256-09	WC-10	BF136155.D	06 Nov 2023 15:27		CG\JU	Ok,M
9	O5257-05	WC-2	BF136156.D	06 Nov 2023 15:57		CG\JU	Ok,M
10	O5256-01	WC-1	BF136157.D	06 Nov 2023 16:28		CG\JU	Ok,M
11	O5253-01	L-1(65FT)(5-10)	BF136158.D	06 Nov 2023 16:57		CG\JU	Ok,M
12	O5253-02	L-6(0-5)	BF136159.D	06 Nov 2023 17:28	Need 10X	CG\JU	Dilution
13	O5256-05	WC-11	BF136160.D	06 Nov 2023 17:59	Need 2X	CG\JU	Dilution
14	O5257-01	WC-6	BF136161.D	06 Nov 2023 18:29	Need 2X	CG\JU	Dilution
15	O5257-05MS	WC-2MS	BF136162.D	06 Nov 2023 18:59		CG\JU	Ok,M
16	O5257-05MSD	WC-2MSD	BF136163.D	06 Nov 2023 19:29		CG\JU	Ok,M
17	O5253-02DL	L-6(0-5)DL	BF136164.D	06 Nov 2023 20:00		CG\JU	Ok,M
18	SSTDCCC040	SSTDCCC040EC	BF136165.D	06 Nov 2023 21:01		CG\JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110723

Review By	yogesh	Review On	11/7/2023 4:17:42 PM
Supervise By	mohammad	Supervise On	11/9/2023 3:24:19 AM
SubDirectory	BF110723	HP Acquire Method	BNA_F
		HP Processing Method	BF110723

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6271 SP6330,SP6331,SP6332,SP6333,SP6334,SP6335,SP6336,SP6337
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6333 S11510 10ul/1000ul sample SP6324

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF136166.D	07 Nov 2023 09:28		CG\JU	Ok
2	SSTDCCC040	SSTDCCC040	BF136167.D	07 Nov 2023 09:58	Need ICAL	CG\JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF136168.D	07 Nov 2023 10:30		CG\JU	Ok
4	SSTDICC005	SSTDICC005	BF136169.D	07 Nov 2023 11:01	Comopunds#32,54,65,77,85 removed from 5 ppm	CG\JU	Ok,M
5	SSTDICC010	SSTDICC010	BF136170.D	07 Nov 2023 11:31	Method is Good For DOD except compound#77	CG\JU	Ok,M
6	SSTDICC020	SSTDICC020	BF136171.D	07 Nov 2023 12:01	Compound #32,54,65 Kept on LR	CG\JU	Ok
7	SSTDICCC040	SSTDICCC040	BF136172.D	07 Nov 2023 12:31	Method is Good For DOD Except Com#77 and good for 625.1 Method	CG\JU	Ok,M
8	SSTDICC050	SSTDICC050	BF136173.D	07 Nov 2023 13:02		CG\JU	Ok,M
9	SSTDICC060	SSTDICC060	BF136174.D	07 Nov 2023 13:33		CG\JU	Ok,M
10	SSTDICC080	SSTDICC080	BF136175.D	07 Nov 2023 14:03		CG\JU	Ok,M
11	SSTDICV040	ICVBF110723	BF136176.D	07 Nov 2023 15:17		CG\JU	Ok,M
12	PB156921BL	PB156921BL	BF136177.D	07 Nov 2023 15:47		CG\JU	Ok
13	PB156921BS	PB156921BS	BF136178.D	07 Nov 2023 16:18		CG\JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM103023

Review By	yogesh	Review On	10/31/2023 4:29:12 AM				
Supervise By	mohammad	Supervise On	10/31/2023 4:55:39 AM				
SubDirectory	BM103023	HP Acquire Method	BNA_M	HP Processing Method	BM103023		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		SP6271 SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		SP6232 S11507 10ul/1000ul sample SP6324					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM042487.D	30 Oct 2023 09:52		MA/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM042488.D	30 Oct 2023 11:04		MA/JU	Ok
3	SSTDICC005	SSTDICC005	BM042489.D	30 Oct 2023 11:40	Comopunds#32,35,41,42,54,56,65,70 removed from 5 ppm	MA/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM042490.D	30 Oct 2023 12:16		MA/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM042491.D	30 Oct 2023 12:52	The Calibration is good for Methods 625.1, 8270 DOD & 8270 NON-DOD Except Benzidine as failed in ICV	MA/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM042492.D	30 Oct 2023 13:29	Compounds#26,32,41,48,53,54,56,57,62,65,91 kept on LR	MA/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM042493.D	30 Oct 2023 14:05		MA/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM042494.D	30 Oct 2023 14:41		MA/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM042495.D	30 Oct 2023 15:18	Compounds#32,73 removed from 80 ppm	MA/JU	Ok,M
10	SSTDICV040	ICVBM103023	BM042496.D	30 Oct 2023 16:23		MA/JU	Ok,M
11	PB156597TB	PB156597TB	BM042497.D	30 Oct 2023 17:00		MA/JU	Ok
12	PB156526BS	PB156526BS	BM042498.D	30 Oct 2023 17:36		MA/JU	Ok,M
13	PB156526BSD	PB156526BSD	BM042499.D	30 Oct 2023 18:13		MA/JU	Ok,M
14	PB156526BL	PB156526BL	BM042500.D	30 Oct 2023 18:49		MA/JU	Ok
15	O4909-01RE	LCOL-12RE	BM042501.D	30 Oct 2023 19:31	Surrogate Fail	MA/JU	Confirms
16	O4958-02RE	EFF-WASTE-WATERR	BM042502.D	30 Oct 2023 20:07	Surrogate Fail	MA/JU	Confirms
17	O5011-01DL	RBR-200027DL	BM042503.D	30 Oct 2023 20:43		MA/JU	Ok



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM103023

Review By	yogesh	Review On	10/31/2023 4:29:12 AM		
Supervise By	mohammad	Supervise On	10/31/2023 4:55:39 AM		
SubDirectory	BM103023	HP Acquire Method	BNA_M	HP Processing Method	BM103023
STD. NAME		STD REF.#			
Tune/Reschk		SP6271			
Initial Calibration Stds		SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236			
CCC		SP6232			
Internal Standard/PEM		S11507 10ul/1000ul sample			
ICV/I.BLK		SP6324			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM111023

Review By	yogesh	Review On	11/10/2023 2:59:37 PM				
Supervise By	mohammad	Supervise On	11/16/2023 1:32:18 AM				
SubDirectory	BM111023	HP Acquire Method	BNA_M	HP Processing Method	BM103023		
STD. NAME		STD REF.#					
Tune/Reschk		SP6271					
Initial Calibration Stds		SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236					
CCC		SP6232					
Internal Standard/PEM		S11528 10ul/1000ul sample					
ICV/I.BLK		SP6324					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM042684.D	10 Nov 2023 10:10		MA/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM042685.D	10 Nov 2023 11:22	CCC failed high for comp. #04,15,19,34,41,42,61,67,68,77,79	MA/JU	Ok,M
3	PB156988BL	PB156988BL	BM042686.D	10 Nov 2023 11:58	Surrogate Fail	MA/JU	Not Ok
4	O5279-04	TP-1	BM042687.D	10 Nov 2023 12:34		MA/JU	Ok
5	O5279-08	MH-1	BM042688.D	10 Nov 2023 13:10	Internal Standard Fail	MA/JU	ReRun
6	O5295-04	TP-2	BM042689.D	10 Nov 2023 13:46	Internal Standard Fail	MA/JU	ReRun
7	O5311-03	TP-1	BM042690.D	10 Nov 2023 14:22	Internal Standard Fail	MA/JU	ReRun
8	O5280-02	MH-1-GW	BM042691.D	10 Nov 2023 14:58	Internal Standard Fail	MA/JU	ReRun
9	O5317-03	RB-21148	BM042692.D	10 Nov 2023 15:34	Need 5X	MA/JU	Dilution
10	O5279-01	TP-1	BM042693.D	10 Nov 2023 16:09		MA/JU	Ok,M
11	O5279-05	MH-1	BM042694.D	10 Nov 2023 16:46		MA/JU	Ok,M
12	O5317-01	208	BM042695.D	10 Nov 2023 17:22		MA/JU	Ok,M
13	O5292-01	CORONA	BM042696.D	10 Nov 2023 17:58	Internal Standard Fail	MA/JU	Ok,M
14	O5291-01	QUEEN-PLAZA	BM042697.D	10 Nov 2023 18:34		MA/JU	Ok,M
15	O5252-01	WASTE	BM042698.D	10 Nov 2023 19:10		MA/JU	Ok
16	O5253-03	L-3(120FT)(0-5)	BM042699.D	10 Nov 2023 19:46	Need Further 10X	MA/JU	Dilution
17	O5253-04	L-3(195FT)(0-5)	BM042700.D	10 Nov 2023 20:22	Need Further 5X	MA/JU	Dilution
18	O5257-01DL	WC-6DL	BM042701.D	10 Nov 2023 20:58		MA/JU	Ok,M



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM111023

Review By	yogesh	Review On	11/10/2023 2:59:37 PM				
Supervise By	mohammad	Supervise On	11/16/2023 1:32:18 AM				
SubDirectory	BM111023	HP Acquire Method	BNA_M	HP Processing Method	BM103023		
STD. NAME	STD REF.#						
Tune/Reschk	SP6271						
Initial Calibration Stds	SP6228,SP6229,SP6230,SP6231,SP6232,SP6233,SP6234,SP6235,SP6236						
CCC	SP6232						
Internal Standard/PEM	S11528 10ul/1000ul sample						
ICV/I.BLK	SP6324						
Surrogate Standard							
MS/MSD Standard							
LCS Standard							
19	O5256-05DL	WC-11DL	BM042702.D	10 Nov 2023 21:35		MA/JU	Ok,M

M : Manual Integration