



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Surrogate Summary

SW-846

SDG No.: O5252

Client: RMJ Environomics, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
O5252-01	WASTE	2-Fluorophenol	150	86.3	58		30 (18)	130 (112)
		Phenol-d6	150	81.2	54		30 (15)	130 (107)
		Nitrobenzene-d5	100	60.6	61		30 (18)	130 (107)
		2-Fluorobiphenyl	100	62.8	63		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	92.7	62		30 (10)	130 (110)
		Terphenyl-d14	100	63.8	64		30 (14)	130 (112)
O5257-05MS	WC-2MS	2-Fluorophenol	150	112	74		30 (18)	130 (112)
		Phenol-d6	150	111	74		30 (15)	130 (107)
		Nitrobenzene-d5	100	86.1	86		30 (18)	130 (107)
		2-Fluorobiphenyl	100	84.6	85		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	117	78		30 (10)	130 (110)
		Terphenyl-d14	100	69.8	70		30 (14)	130 (112)
O5257-05MSD	WC-2MSD	2-Fluorophenol	150	120	80		30 (18)	130 (112)
		Phenol-d6	150	118	79		30 (15)	130 (107)
		Nitrobenzene-d5	100	90.3	90		30 (18)	130 (107)
		2-Fluorobiphenyl	100	90.4	90		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	127	85		30 (10)	130 (110)
		Terphenyl-d14	100	75.3	75		30 (14)	130 (112)
PB156921BL	PB156921BL	2-Fluorophenol	150	126	84		30 (18)	130 (112)
		Phenol-d6	150	125	84		30 (15)	130 (107)
		Nitrobenzene-d5	100	86.6	87		30 (18)	130 (107)
		2-Fluorobiphenyl	100	85.1	85		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	131	87		30 (10)	130 (110)
		Terphenyl-d14	100	84.9	85		30 (14)	130 (112)
PB156921BS	PB156921BS	2-Fluorophenol	150	127	85		30 (18)	130 (112)
		Phenol-d6	150	124	83		30 (15)	130 (107)
		Nitrobenzene-d5	100	82.1	82		30 (18)	130 (107)
		2-Fluorobiphenyl	100	78.8	79		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	132	88		30 (10)	130 (110)
		Terphenyl-d14	100	84.2	84		30 (14)	130 (112)