



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/06/2023

Lab File ID: BF136149.D Time Analyzed: 12:01

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	106195	6.822	409921	8.10	211105	9.86
UPPER LIMIT	212390	7.322	819842	8.598	422210	10.363
LOWER LIMIT	53097.5	6.322	204961	7.598	105553	9.363
EPA SAMPLE NO.						
01 WC-2MS	84798	6.82	326584	8.10	167090	9.86
02 WC-2MSD	84667	6.82	331240	8.10	166345	9.86

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.