



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252
EPA Sample No.: SSTDCCC040 Date Analyzed: 11/10/2023
Lab File ID: BM042685.D Time Analyzed: 11:22
Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	72991	7.845	331358	10.66	231375	14.50
UPPER LIMIT	145982	8.345	662716	11.163	462750	15.004
LOWER LIMIT	36495.5	7.345	165679	10.163	115688	14.004
EPA SAMPLE NO.						
01 WASTE	62722	7.85	245261	10.66	160144	14.50

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.